



## KERALA EXCISE DEPARTMENT

EXC/3585/2022-XGBCommissionerate of Excise  
Thiruvananthapuram,

Dated:16-09-2022

e-mail Id: eoffice.exc@kerala.gov.in

Phone No:0471-2322825

Sealed quotations are invited for the supply of **217 Numbers of Multi-Drug Detection Kit(Oral fluid)** for the use of Excise Department. The **Multi- Drug Multi-Line kit (Oral fluid)** are to be supplied at Excise Head Quarters, Thiruvananthapuram. The specification, terms and the conditions of the **Multi-Drug Detection Kit(Oral fluid)** is appended herewith. Intending vendors of reputed, recognised and qualified makers of these items may quote their offers separately. The quotations should contain price of the item including taxes, guarantee/warranty, details any special offers from the firm etc. The sealed envelopes containing the quotation should bear the superscription **Multi-Drug Detection Kit(Oral fluid)** and should be addressed to the Excise Commissioner. Intending parties shall submit the quotations on their own letter heads with full postal and e-mail address with all conditions regarding supply of article. Payment will be made only after the satisfactory verification of the supplied items. Advance payment will not be allowed. The maximum period required for delivery of the item should also be mentioned. The **Multi-Drug Detection Kit(Oral fluid)** should be supplied at the Excise Head quarters by the firms. Last date for receipt of quotations is ~~28/09/2022~~ **28/09/2022 at 11:59AM**. Late quotations will not be entertained at any cost. The quotations will be opened at ~~28/09/2022~~ **28/09/2022, 1PM** in the presence of such of the tenderer or their authorized representatives who may be present at that time. One each sample shall be submitted along with the quotation.

The undersigned is competent to cancel, modify or insert any conditions besides the above and reserve the right to cancel the proceedings without assigning any reason.

Signed by

Rajeev D.

Date: 16-09-2022 10:22:36

Copy to:-

Yours Faithfully,

RAJEEV D.

ADDITIONAL EXCISE COMMISSIONER

For Commissioner of Excise.

I/246474/2022

- 1 Nodal Officer, IT cell, Excise Headquarters- For publishing in Kerala Excise website
- 2 Notice Board
- 3 Authorized suppliers. Specifications:
  - 1 Instrument/Kit should have quantitative/Semi quantitative result interpretation
  - 2 Instrument should be used for In-Vitro diagnostics purpose
  - 3 Instrument/Kit should have CE Certification.
  - 4 Instrument/Kit should be based on the principle Biochemistry/Competitive binding assay. 5. Instrument/Kit should have control facility for quality.
  - 5 Oral fluid/Saliva samples should be able to test.
  - 6 The system shall be able to measure the following the parameters
    - a. Cocaine (COC)
    - b. Methamphetamine (MET)
    - c. Phencyclidine (MET)
    - d. Marijuana (THC)
    - e. Opiate (OPI)
    - f. Amphetamine (AMP)